



### Quality Division Use Only

Scientific Area Committee (SAC):  Subcommittee:

Standard/Guidelines Title

Date added to OSAC Registry

Can the Standard/Guidelines be followed in totality?

Please explain if the Standard/Guidelines and the applicable clauses that cannot be followed in their totality:  
N/A

Compliance plan to achieve conformance to Standard/Guidelines:  
Since the standard was not technical in nature, the incorporation of these requirements into the Toxicology procedure did not require modification to analytical processes nor further validation.

Declaration of Compliance to OSAC Registry Standard/Guidelines:  
As of October 5, 2018, the Toxicology Section is compliant with the requirements listed in "Standard Practices for Measurement Traceability in Forensic Toxicology". The required metrological traceability and calibration schedule for analytical equipment (balances, reference standards, pipettes, diluters and thermometers) had already been implemented in the section to meet conformance with accreditation requirements. On October 5, 2018 the Toxicology Section incorporated the frequency of calibration and acceptability/tolerance specifications for volumetric glassware from this standard into their sectional procedure (v3.3) under section 9.14.

Technical Leader:

Date:

Section Manager:

Date:

Division Director:

Date:

Quality Director:

Date: